



Acceptable wavelength band width for HPLC detectors when performing an HPLC Performance Qualification - Tech Information

3nm is the normal setting for wavelength bandwidth for HPLC Detectors. 5nm may be considered too wide and 1nm is too narrow for PQ.

Click [HERE](#) for MicroSolv PQ and HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 02-05-2024