



Solvents and solutions that should be avoided when using PEEK filters fittings or tubing – Tech Information

Solvents and chemicals that should be avoided when using PEEK filters, tubing and fittings include:

- Nitric acid
- Sulfuric acid
- Halogenated Acids, including hydrofluoric, hydrobromic, and hydroiodic acids*
- Pure Halogenated gases
- Methylene chloride
- THF (*Tetrahydrofuran*)
- DMSO (*Dimethyl Sulfoxide*)

**Note that you can use Hydrochloric Acid, all other HPLC grade solvents and HPLC buffers with most applications of PEEK Tubing and Fittings.*

[More information about PEEK tubing](#)

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com