


How do retention and efficiency of asymmetric dimethylarginine ADMA compare using 4um vs 2.0 Diamond Hydride columns – AppNote

ADMA can be retained using an ANP gradient method with the Cogent Diamond Hydride™ HPLC column.

Retention times for ADMA differed only slightly between the 4um and 2.0™ stationary phases. Efficiency however was notably higher when using the 2.0™ phase. This can be readily observed from the greater peak height using the 2.0™ column in the chromatogram overlay comparison shown below:



Method Conditions:

Column dimensions: 2.1 x 50mm (both)

Flow rate: 0.3mL/min

A: DI H₂O + 0.1% formic acid

B: Acetonitrile + 0.1% formic acid

time (min)	%B
0	90
5	30
8	30
10	90

post time 4 min



For more information: [Cogent Diamond Hydride Product Page](#)

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com