



Capillary conditioning in capillary electrophoresis – How To

New capillaries should be conditioned before you use them.

1. Filter all solutions to be used on your capillary through a .45 μ or 0.2 μ syringe filter.
2. Treat the capillaries by running 0.1N NaOH solution through them on your CE unit.
3. Rinse the capillaries with CEwater or HPCE grade water for 5 minutes.
4. Final step is to run your buffer through it for 10 minutes before injection.

This procedure should be followed each day before new or old capillaries are used.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com