



Changing capillary dimensions in capillary electrophoresis – Tech Information

When you change the ID of your capillary in your CZE system, the following changes can be expected. It is important to be aware of these effects.

Joule Heating: An increase ID or length of your capillary will increase the amount of heat due to Joule Heating. Smaller ID and length of capillaries are utilized to more efficiently dissipate Joule Heating.

Resolution:

- 1.A reduction in capillary ID reduces detector sensitivity to due a smaller path length.
- 2.A reduction in capillary ID reduces the sample load requirements.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-08-2024