

## Can I use Silica C columns in a Normal Phase gradient method – $\ensuremath{\mathsf{FAQ}}$

## Yes!

Although this is not easy with ordinary silica columns due to solvent de-mixing & water uptake, the Cogent Silica-C<sup>™</sup> column is not limited by this and is easy to use even with fast gradients.



Printed from the Chrom Resource Center **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435 Email: customers@mtc-usa.com Website: www.mtc-usa.com Date: 05-18-2024