



# Acceptable wavelength band width for HPLC detectors when performing an HPLC Performance Qualification - Tech Information

**3nm is the normal setting for wavelength bandwidth for HPLC Detectors.** 5nm may be considered too wide and 1nm is too narrow for PQ.

Click [HERE](#) for MicroSolv PQ and HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-17-2024