

What is the difference between a union and an adapter in HPLC – $\ensuremath{\mathsf{FAQ}}$

The biggest difference is the price where a union is much less expensive than an adapter.

For performance considerations, a union will make a connection between tubing and fittings that have the same thread size and thread count configurations. An adapter will connect two different thread sizes and thread count.

Click *HERE* for HPLC fittings ordering information and pictures.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435 Email: customers@mtc-usa.com Website: www.mtc-usa.com