



What are the terms such as high flow rate in Controlled Flow CE capillaries based - FAQ

The terms **High**, **Medium** and **Low Flow** are nominal and based on a relative flow when compared to bare fused silica capillaries.

The Controlled Flow capillary always is slower than bare silica but the high rate is the highest of the Controlled Flow capillaries.

A set of conditions on a bare silica column that provides a migration time of 3.2 minutes for a neutral marker will provide a migration time of 4.3 minutes for High, 6.5 minutes for Medium and 8.2 minutes for Low when using Controlled Flow capillaries with that same marker.

Click [HERE](#) to view Zero and Controlled flow capillaries for **CE ordering information**.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-11-2024