



## Is the conditioning of the Controlled Flow and Zero CE capillaries the same as with a bare fused silica capillary – FAQ

**No. Inside the front cover of the capillary case is a special set of conditioning steps that must be taken.**

It is important that these conditions be followed. You can also see these conditions on our website by clicking on this link: [Controlled flow capillary conditioning steps](#).

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-07-2024