



Can I use the knitted reactor coils KRC with my HPLC fluorescence detector – FAQ

The only restriction on whether or not you can use the knitted reactor coils KRC is the inlet tubing of your detector. The absolute minimum ID can be 0.008". It is preferred that the tubing ID be 0.01". Anything smaller than 0.008" is not compatible.

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com