



Will I see any differences in my HPLC chromatography if I change to AQ syringe filters – FAQ

If you use proper filtering techniques and a proper filter you should not see any differences in your chromatography, CE or UV testing. Syringe filters only protect your equipment and columns and never participate in separations or SPE.

Click [HERE](#) for syringe filters ordering information and pictures.

Attachment

MicroSolv filters equivalency study pdf [Download File](#)

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-07-2024