



## Are the MicroSolv AQ brand of syringe filter membranes supported by extra membranes in the plastic housing - FAQ

**The PTFE Teflon syringe filter membranes are backed with an inert polypropylene membrane for support.**

All the others syringe filters including Nylon, Regenerated Cellulose, Cellulose Acetate, PVDF etal do not require additional support and therefore have no additional membranes.

Click [HERE](#) for ordering information and pictures of AQ brand syringe filters

### **Attachment**

**MicroSolv Filters Equivalency Study pdf** [Download File](#)

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-06-2024