




## Tip for Promoting Analyte Ionization with LCMS – AppNote

### Analyte Ionization using L-Tyrosine

Figures A and B show the Peak for L-Tyrosine obtained with two LCMS Methods. The analysis shown in *Figure B* was performed with a stronger ANP Solvent (DI Water) as component A of the Mobile Phase. The obtained Retention time was lower than in *Figure A*, as was the compound ionization.

An eluent containing 2-Propanol is a better choice for the analysis of this compound, since the ionization of L-Tyrosine is greater. In addition, the Solvents used in *Figure A* are recommended when biological samples are analyzed. As such, the Method can be directly transferred to the analysis of L-Tyrosine in plasma samples for monitoring of patients with Phenylketonuria.





L-Tyrosine 182.0812 m/z

## Method Conditions

**Column:** Cogent Diamond Hydride™, 4μm, 100Å

**Catalog No.:** 70000-15P-2

**Dimensions:** 2.1 x 150mm

**Mobile Phase:**

*Figure A*

A: 50% DI Water / 50% 2-Propanol / 0.1% Formic Acid v/v

B: Acetonitrile / 0.1% Formic Acid v/v

*Figure B*

A: DI Water / 0.1% Formic Acid v/v

B: Acetonitrile / 0.1% Formic Acid v/v

**Gradient:**

Time (minutes)	%B
0	90
4	30
6	30
7	90

**Post Time:** 3 minutes

**Injection vol.:** 1μL

**Flow rate:** 0.4mL / minute

**Detection:** ESI - POS - Agilent 6210 MSD TOF Mass Spectrometer

**Sample Preparation:** 500mg L-Tyrosine capsule contents were dissolved in 500mL of 50:50 Water / Acetonitrile.

The solution was shaken for 20 minutes and filtered through a disposable 0.45μm Nylon Syringe Filter (MicroSolv Tech Corp.). The sample for injection was diluted 1:100 with a 50:50 Solvent A / Solvent B mixture.

**t<sub>0</sub>:** 0.9 minutes

**Note:** ANP Methods have an inherent advantage in LCMS over Reversed Phase Methods in terms of Sensitivity. This is because Acetonitrile is more volatile than Water and hence more readily removed in LCMS, and ANP Methods typically use a higher percentage of Acetonitrile in the Mobile Phase compared to Reversed Phase Methods.

Printed from the Chrom Resource Center