



## Are nylon membrane and syringe filters compatible with chloroform - FAQ

**All nylon filters are not recommended for use with halogenated solvents such as chloroform or with strong acids (1N and stronger).**

For these liquids, you should consider using PTFE filters.

Click [HERE](#) for ordering information and pictures of AQ brand of syringe filters.

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)