



Can I sterilize screw caps for autosampler vials that have silicone rubber with PTFE septa - FAQ

Yes, if you use heat / steam or gamma irradiation, the AQ™, MicroSolv™ and the BASIK™ brands of 8mm, 9mm and 10mm screw caps can be sterilized or autoclaved for up to 15 minutes.

Click [HERE](#) for ordering information and pictures of autosampler vials

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-15-2024