

Solving baseline problems in ANP methods with the Cogent Diamond Hydride column – Tips & Suggestions

When using a gradient method with the Cogent Diamond Hydride™ column in ANP HPLC, and an inconsistent baseline is observed, following are some possible causes of the issue along with suggestions to fix them.

1. Is the mobile phase filtered before use on the instrument? A 0.45µm nylon membrane filter with vacuum filtration is recommended. This is especially important for mobile phases that contain dissolved solids such as ammonium acetate. It is good practice to filter any mobile phase though.
2. How many runs were tried? When ammonium formate / acetate is involved in the mobile phase, it sometimes can take three runs or so to get a reproducible baseline in Aqueous Normal Phase HPLC with the [Cogent Diamond Hydride™ column](#).
3. How much post time is used in the gradient? Please allow a sufficient post time to insure the mobile phase is fully equilibrated with the starting gradient solvent conditions before each injection.
4. Try injecting multiple blanks and see if the baseline becomes more consistent. The inconsistent baseline may be due to carryover of strongly retained material in the sample that slowly elute in subsequent runs. If this is the case, your method will need to incorporate stronger eluting conditions at the end of the gradient program to wash these compounds out so that they do not carry over into the next run.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 06-05-2024