


# Caffeine in Coffee Analyzed by HPLC- AppNote

# **Unique Retention Mode Affords Superior Specificity**

Although Caffeine retains well in Reversed Phase, it is found to be difficult to obtain a well-resolved Peak free of interference from other matrix peaks with a complex sample such as coffee. In this Method, most of the matrix peaks elute near the void volume and do not interfere with the Caffeine Peak, which is well-resolved from the others.

Complex matrices can also adversely affect run-to-run repeatability due to compounds that do not elute from the Column and change the chromatography. Here the data shows no sign of contaminant build-up on the Column, as the run-to-run overlays show. The lot-to-lot reproducibility is good as well. Finally, the Method conditions are LCMS compatible.





## **Method Conditions**

**Column:** Cogent Diamond Hydride<sup>™</sup>, 4μm, 100Å

**Catalog No.:** 70000-7.5P **Dimensions:** 4.6 x 75mm

**Mobile Phase:** 

A: DI Water / 0.1% Formic Acid (v/v)
B: Acetonitrile / 0.1% Formic Acid (v/v)

### **Gradient:**

Time (minutes)	%B
0	98
2	98
7	50
8	98

Post Time: 3 minutes
Injection vol.: 1µL

Flow rate: 1.0mL / minute **Detection:** UV @ 275nm

**Sample Preparation:** Commercially available ground coffee was brewed and filtered with a 0.45µm Nylon Syringe Filter (MicroSolv Tech Corp.). It was then diluted 1:10 with a diluent of 50:50 Solvent A / Solvent B. The Caffeine Peak identity was confirmed with a USP reference standard.

**to:** 0.9 minutes

**Note:** Caffeine is a Xanthine Alkaloid found in the coffee plant, the tea bush, the kola nut, and other plants. It is the most commonly consumed psychoactive drug in the world.



#### Attachment

No 223 Caffeine in Coffee Analyzed by HPLC pdf 0.4 Mb Download File

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation** 

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Website: www.mtc-usa.com

Email: customers@mtc-usa.com

Date: 05-17-2024