


# Streptomycin Analyzed by HPLC - AppNote

## **Retention of a Highly Polar Antibiotic**

Streptomycin is very polar and difficult to retain with Reversed Phase methods using conventional Columns. In this Method, retention is achieved with Separation of the Streptomycin from two impurities found in the Standard Solution. Furthermore, the numerous amine groups do not cause tailing of Peak shape as with standard silica-based Columns.

Data from two Column lots shown in the figure illustrates the reproducibility of the Method and it robustness.



#### **Peaks:**

- 1. Impurity
- 2. Streptomycin



3. Impurity

### **Method Conditions**

Column: Cogent Diamond Hydride™, 4µm, 100A

**Catalog No.:** 70000-7.5P **Dimensions:** 4.6 x 75mm

**Mobile Phase:** 

A: DI Water / 0.1% Trifluoroacetic Acid (v/v)
B: Acetonitrile/ 0.1% Trifluoroacetic Acid (v/v)

#### Gradient:

Time (minutes)	%B
0	95
1	95
6	40
7	95

Post Time: 3 minutes
Injection Vol.: 1µL

Flow Rate: 1.0mL / minute Detection: UV @ 205nm

Sample Preparation: 1.0mg / mL Streptomycin Sulfate reference standard in Solvent A diluent.

to: 0.9 minutes

**Note:** Streptomycin is an aminoglycoside antibiotic derived from the Actinobacterium Streptomyces Griseus. Streptomycin was isolated in 1943 and was the first antibiotic to treat tuberculosis.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation** 

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 03-05-2024