



## Is stainless steel or titanium better for mobile phase inlet filters - FAQ

Stainless steel is suitable for most applications but titanium is recommended for analyses involving biological analytes and matrices.

[Stainless steel inlet filter product page](#)

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-05-2024