

Specifications of different grades of stainless steel used in HPLC columns

Cogent™ HPLC stainless steel column hardware is T1 grade, AISI 316 LC type. HPLC system tubing is AISI 316 (highlighted in green in the chart below, T2 grade. *We also offer fully coated stainless steel column hardware and PEEK column hardware to eliminate sample interaction with the surface sites of stainless steel.*

Elemental assays of the various stainless steel grades are given in the following table:



Ti Grade 2

Carbon	0.1 max
Hydrogen	0.015 max
Iron	0.3 max
Nitrogen	0.03 max
Oxygen	0.25 max
Titanium	Balance



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-11-2024