



Vials are listed as 12 x 32mm but my autosampler accepts a maximum diameter of 11.8mm will they fit –
FAQ

12 x 32mm OD, 2ml vials are listed as 12mm diameter for standardization purposes within the industry. This is a nominal value only (name only) and the exact value is suitable for use with your autosampler.

The outer diameter is actually 11.8mm for most 12x32mm vials.

Click [HERE](#) for autosampler vial ordering information and pictures.

AUTOSAMPLER
VIALS AND CAPS

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-19-2024