



How much volume can the 47mm HPLC membrane filters process – FAQ

In most cases, one of these filters can be used until it is blocked with particulate matter at which time, the flow through the membrane will cease.

In most applications you will use the filters for removing small amounts of particulate matter from your HPLC mobile phase before introducing them to the system. The filter should be suitable for a wide range of solvent volume in this case, since there will be very little particulate matter present.

In other situations such as filtering turbid Liquid Samples (e.g. milk, creams, lotions), the Filter will become blocked much quicker. Consider Centrifuging the liquid First before Filtering in situations like this.

[47mm Membrane Filter Product Page](#)

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 03-05-2024