

Can I use the Cogent Diamond Hydride column without mobile phase additives like formic acid – FAQ

Yes you can. Using only acetonitrile and water with no mobile phase additives is fully compatible with the column and will not damage it.

If you find the peak shape is not as good without formic acid, try using a steeper gradient to sharpen it. Some compounds' retention and / or peak shape may be significantly affected by the presence of a mobile phase **additive** and others may not. It depends on various factors pertaining to the **analyte**, the stationary phase, and so on.



Click [HERE](#) for Cogent Diamond Hydride™ ordering information

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-15-2024