



## What is the inside diameter of the opening of your Luer Lock syringes catalog number 58901-S - FAQ

The inner diameter (ID) of the opening of the Chrom Syringes has been measured to be approximately 2.1–2.2mm.



[Syringes for Chromatography Ordering Information](#)

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-16-2024