

Which type of vial do you recommend to use with the U-2D Micro Sample Management inserts – FAQ

We recommend our line of AQ™ brand 9mm screw top or snap top vials which are manufactured to mate precisely the vial's inner dimensions and the inserts length and outer dimensions, offering maximum performance unmatched by anyone for quality control and the specs required for consistent performance.

Other brands of standard wide-opening 9mm 12x32mm screw top autosampler vials or wide opening snap top 12x32mm vial may be used but we since many brands do not consistently control the bottom height of their vials correctly, we cannot endorse using these vials. For optimal results, use the AQ™ brand of vials.



Drop the Insert into a Vial

Click [HERE](#) for U-2D Micro-Sample Management System Ordering Information

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-19-2024