



Are the AQ brand of nylon syringe filters suitable for sterile filtration – FAQ

The AQ™ syringe filters are not designed for sterile filtration.

Their primary intended function is to remove particulate matter from a sample to be injected into and protect an HPLC instrument and the HPLC column.

Click [HERE](#) for ordering information and pictures of AQ™ brand syringe filters

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 04-05-2024