

## Determining HPLC method accuracy – How To

The accuracy of an HPLC method is the closeness of the measured value to the true value for the sample.

To determine the accuracy of a proposed method, different levels of the analyte concentrations: lower concentration (LC, 80%), intermediate concentration (IC, 100%) and higher concentration (HC, 120%) must be prepared from independent stock solutions and analyzed (n=10).

Accuracy is assessed as the percentage relative error and mean %recovery. To provide an additional support to the accuracy of the developed assay method, standard addition method should be employed, which involves the addition of different concentrations of the analyte (for example: 10, 20 and 30 µg/mL) to a known pre-analyzed sample and the total concentration should be determined using the proposed methods (n=10).

The %recovery of the added analyte should be calculated as:

$$\%recovery = [(Ct - Cs) / Ca] \times 100$$

where Ct is the total analyte concentration measured after standard addition; Cs, analyte concentration in the sample; Ca, analyte concentration added to sample.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 04-05-2024