




Oxybenzone and Octinoxate Analyzed by HPLC - AppNote

Separation of Two APIs in Chap Stick® Extract

This Method shows how two common ingredients found in sunscreens and lip balms can be separated. The two compounds are very hydrophobic, so a Mobile Phase Gradient with significant organic content was used in order to avoid excessive retention.

Likewise, a highly organic diluent should be used to adequately extract the compounds from the lip balm material. The *figure* below shows an overlay of two runs from different Column lots, demonstrating the robustness and precision of the Method.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Peaks:

1. Oxybenzone
2. Octinoxate

Method Conditions

Column: Cogent Bidentate C18™, 4µm, 100Å

Catalog No.: 40018-75P

Dimensions: 4.6 x 75mm

Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v)

B: Acetonitrile / 0.1% Formic Acid (v/v)

Gradient:

Time (minutes)	%B
0	60
1	60
4	100
6	100
7	60

Temperature: 40°C

Post Time: 3 minutes

Injection vol.: 2µL

Flow rate: 1.0mL / minute

Detection: UV @ 288nm

Sample Preparation: 250mg of ChapStick containing 7.5% Octinoxate and 3.5% Oxybenzone was weighed in a 25mL volumetric flask and a portion of 90:10 Acetonitrile / DI Water was added. It was sonicated for 30 minutes and diluted to mark. Then a portion was filtered with a 0.45µm Nylon Syringe Filter (MicroSolv Tech Corp.). The filtrate was diluted 1:10 for injections.

t₀: 0.9 minutes

Note: These two compounds exhibit strong absorbance in the UV range, hence the wavelength of 288nm that was chosen. This behavior accounts for their use as sunscreen agents; UV radiation that is absorbed by the compounds is then prevented from reaching the skin and causing damage.



Attachment

No 216 Oxybenzone and Octinoxate Analyzed by HPLC pdf 0.3 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com