


Sertraline Tablet Analyzed with HPLC - AppNote

Retention of Hydrophobic and Hydrophilic Analytes with Zoloft®

Sertraline has a logP of 2.9, but retains well when using Aqueous Normal Phase (*ANP*) HPLC. A polar test solute, Phenylglycine, was added to demonstrate how both nonpolar and polar compounds can be retained in the same Method. This would be important when looking for impurities or degradants in the formulation.

Furthermore, Sertraline's amine group can produce problematic tailing with some Reversed Phase columns, but the Peak Symmetry obtained with this Method is very good. Data from two column lots is presented to illustrate Method Robustness and lot-to-lot consistency.



Phenylglycine

Sertraline

d from the Chrom Resource Center Copyright 2024, All Rights Apply

Peaks:

1. Phenylglycine

2. Sertraline

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com



Method Conditions

Column: Cogent Diamond Hydride[™], 4μm, 100Å

Catalog No.: 70000-7.5P **Dimensions**: 4.6 x 75mm

Mobile Phase:

A: DI Water with 0.1% Formic Acid (v/v)
B: Acetonitrile with 0.1% Formic Acid (v/v)

Gradient:

Time (minutes)	%B
0	95
1	95
6	50
7	95

Post Time: 3 minutes
Injection vol.: 1µL

Flow rate: 1.0mL / minute Detection: UV @ 254nm

Sample Preparation: 25mg strength Sertraline (Zoloft®) tablet was ground and added to a 25mL volumetric flask. A portion of 50:50 Solvent A / Solvent B diluent was added and the flask was sonicated 10 minutes. It was then diluted to mark and mixed. A portion was filtered with a 0.45µm Nylon Syringe Filter (MicroSolv Tech Corp.) and mixed 1:1 with a 1mg / mL Phenylglycine solution.

to: 0.9 minutes

Note: Sertraline is a selective Serotonin reuptake inhibitor used to treat major depressive disorder. It is widely prescribed and sold as Zoloft® from Pfizer.



Attachment

No 246 Sertraline Tablet Analyzed with HPLC pdf 0.3 Mb Download File

Printed from the Chrom Resource Center
Copyright 2024, All Rights Apply
MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com