


Tizanidine HCl HPLC Method transferred to near UHPLC – AppNote

Use of near UHPLC column for improved results

This challenging compound has several amine groups and can be problematic to analyze by HPLC. Here a good peak shape is obtained using the Cogent Diamond Hydride 2.0 Column. The efficiency is higher on the 2.2 μ m phase compared to a standard 4 μ m Column, leading to increased sensitivity. The method conditions are compatible with LCMS as well.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply


MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com



Tizanidine

Peak:

Tizanidine HCl

Method Conditions

Columns:

Fig. A: Cogent Diamond Hydride 2.0™, 120Å

Fig. B: Cogent Diamond Hydride™, 4µm, 100Å

Catalog Nos.:

Fig. A: [70200-05P-2](#);

Fig. B: [70000-05P-2](#)

Dimensions:

Fig. A: 2.1 x 50 mm

Fig. B: 2.1 x 50 mm

Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v)

B: Acetonitrile / 0.1% Formic Acid (v/v)

Gradient:

Time (minutes)	%B
0	95
1	95
6	40
7	95

Post time: 3 minutes

Injection vol.: 0.2 µL

Flow rate: 0.3mL/minutes

Detection: UV @ 230 nm

Sample: 4mg strength Tizanidine HCl tablet was ground and weighed in a 10mL Volumetric Flask. A portion of 50/50 Solvent A / Solvent B diluent was added and the flask was sonicated 10 minutes. It was then diluted to mark and filtered with a 0.45µm Nylon Syringe Filter (MicroSolv Tech Corp.).

t₀: 0.6 minutes

Note: Tizanidine is a centrally acting A2-Adren- Ergic agonist used to treat spasms, cramping, tightness of muscles,

Printed from the Chrom Resource Center
Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

and related conditions. It is available under the trade name Zanaflex® as well as generic versions.



Attachment

Tizanidine HCl Method Transfer pdf [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com