



## LSD and Codeine - AppNote

### LCMS Method for Spiked Urine Sample

LSD and codeine may be present in Urine samples and methods are needed for their detection in these matrices. The extracted Ion Chromatograms (EICs) for both analytes are shown in the figure and illustrate excellent chromatographic as well as Mass Spectral Separation. The Method in this AppNote produces high Efficiency for both Peaks.



#### Peaks:

1. LSD 324.207 m/z [M + H]<sup>+</sup>

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

2. Codeine 300.1594 m/z [M + H]<sup>+</sup>

## Method Conditions

**Column:** Cogent Diamond Hydride 2.0™, 2.2µm, 120Å

**Catalog No.:** 70200-05P-2

**Dimensions:** 2.1 x 50mm

### Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v)

B: Acetone / 0.1% Formic Acid (v/v)

### Gradient:

Time (Minutes)	%B
0	85
3	30
5	30
6	85

**Flow rate:** 0.4 mL/minute

**Detection:** ESI - POS - Agilent 6210 MSD TOF Mass Spectrometer

**Injection vol.:** 1µL

**Sample Preparation:** Urine spiked with 5ppm LSD and Codeine standard solutions

**Note:** Lysergic Acid Diethylamide (LSD) was synthesized by Swiss chemist Albert Hofmann while researching Ergot Alkaloid Derivatives. It is highly potent and produces significant psychedelic effects when ingested. It is currently a Schedule I drug in the U.S. Codeine is an opiate used for analgesic, antitussive, and other effects. It may be obtained from over-the-counter formulations like Cough Syrup and therefore has potential for abuse.



### Attachment

No 326 LSD and Codeine pdf 0.3 Mb [Download File](#)