

Capillary storage in capillary electrophoresis - How To

When storing a capillary after use, follow these simple instructions to increase the capillary life.

- 1. Filter all solutions to be used on your capillary with .45 μ or .2 μ filter.
- 2. Rinse the capillary with .1M NaOH solution for 5 minutes.
- 3. Rinse the capillary with CEwater or CE grade water for 5 minutes.
- 4.If possible, blow nitrogen or helium through the capillary prior to storage.

Printed from the Chrom Resource Center Copyright 2024, All Rights Apply MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com