



## Changing Conductivity in Capillary Electrophoresis.

When you change the Conductivity of your CZE method, the following changes can be expected.

**Joule Heating:** Reducing or lowering the conductivity will reduce the amount of Joule Heating in your capillary.

**Current:** A lower conductivity results in a lower Current.

**Resolution:** When conductivity is reduced sample loading can be adversely effected.

**Migration Time:** A greater difference in Conductivity between the sample buffer (sample zone) and your run buffer will speed migration time. This is called Stacking.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-18-2024