

Viscosity changes will affect these parameters in capillary electrophoresis – Tips & Suggestions

When the viscosity of your CZE method changes, the following can be expected. Changing anyone of these parameters in CZE can reduce your coefficient of variation therefore it is important to control your viscosity accurately.

Electro Osmotic Flow (EOF): A lower viscosity will increase the EOF in your CZE method.

Analyte to Wall Interaction: A non-homogeneous viscosity will cause analyte to wall effects that can result in non-reproducible migration times and Peak Tailing.

Electrophoretic Mobility: A decrease in viscosity can cause an increase in electrophoretic mobility

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-26-2024