
Can the autosampler vials be used in a centrifuge? What is the maximum centrifugal force that can be used?

Autosampler vials are not designed to be centrifuged so we cannot support it with the exception of the MRQ & Max **Recovery** vials due to their thicker walls and “single piece” construction (AQ brand, RSA or RSA-Pro). They can withstand g forces up to 6,000 x g when used in a compatible centrifuge and adapters.

[The Vial Centrifuge Product Page](#)