



How are MicroSolv brand vials packaged to ensure no contamination – FAQ

All MicroSolv brand vials that carry a CleanPack™ label on the front side of the polypropylene box have been packed in a certified clean room, class 10,000 after having passed the Lehrer oven at approximately 600°C.

You consumer can rely on clean, uncontaminated vials for a correct analysis.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com