



Will phosphates precipitate in an HPLC column – FAQ

Phosphoric acid will not precipitate in the columns. Sodium phosphate (which is not recommended) but which is often used as a buffer, will precipitate in the columns and should be avoided as a storage solvent for all HPLC columns.

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com