

---

## Sample diluents should match your starting mobile phase - Tips & Suggestions

For best results, it is always best to try to dissolve your samples in as close as possible to your isocratic **eluent** or the start of your chosen gradient.

If this is not possible gradually increase the percentage of the solvent in your **eluent** in which the target is most soluble.

