



Can I use Silica C columns in a Normal Phase gradient method – FAQ

Yes!

Although this is not easy with ordinary silica columns due to solvent de-mixing & water uptake, the Cogent Silica-C™ column is not limited by this and is easy to use even with fast gradients.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-17-2024