



Can I use Silica C columns in a Normal Phase gradient method - FAQ

Yes!

Although this is not easy with ordinary silica columns due to solvent de-mixing & water uptake, the Cogent Silica-C™ column is not limited by this and is easy to use even with fast gradients.



MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com