



## Why are the Cogent Silica-C columns more stable than other silica columns - FAQ

We have done extensive testing in our labs and in real world use around the world and have proven that the Cogent Silica-C™ columns are stable even in the presence of water due to the silica hydride's slightly hydrophobic surface.

However, these columns are designed and specified for Normal Phase and **Aqueous Normal Phase** ANP HPLC and not for Reversed Phase. But the only real limitation is your creativity.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024