



Retention time varies from instrument to instrument – Tips & Suggestions

To duplicate the same retention times for analyzed compounds from other sources of methods, as in application notes, gradient conditions have to be adjusted according to the instrument used for parameters such as dwell volume, pump accuracy etc.

Also, each instrument has different plumbing (*dimensions of the tubing*), so obtaining *exactly* the same retention times may be virtually impossible. Therefore, using alpha values or relative retention times is a better way to determine precision and reproducibility.

This can be especially significant when attempting to duplicate application notes from our online Resource Center and website.



MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com