

Use a dry 5ml volumetric flask and a calibrated stopwatch to confirm the flow accuracy at 1.0 mL/minute.

Note that this flow rate check is not meant to replace the typically more extensive operational qualification checks at several flow rates. Rather it is to document the suitability of the pump to perform various PQ Test protocols described in the PQ and HSQ Kits $^{\mathsf{TM}}$.

Click HERE for PQ Kit and HSQ Kit ordering information.

Printed from the Chrom Resource Center
Copyright 2024, All Rights Apply
MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com Website: www.mtc-usa.com