
Can I measure the extra column volume of my HPLC system and compare it to extra column volume of another instrument?

Yes, that is one of the added features of the MicroSolv HSQ Kit™. Using the Rs Test Mixture, and the Excel template, using sophisticated algorithms, a very good estimation of your instruments extra column volume is easily determined.

To compare it to another instrument, you must use the same column under the exact same conditions and run the same test. It is very valuable information for labs that transfer or move methods from one instrument to another.