



## Differences between the linearity solutions in the HSQ qualification kit – Tech Information

They are all different and with precise concentrations of **caffeine** that allows you to easily and accurately determine the **linearity** of your HPLC UV detector.

L3 solution is the only **linearity** solution that has a **uracil** peak present. This peak is used as a pump flow stability marker and has no effect on caffeine or **linearity** tests.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-17-2024