

---

## What are the differences between the linearity solutions in the HSQ Kit for HPLC?

They are all different, precise concentrations of **Caffeine** that allow you to easily and accurately determine the **linearity** of your HPLC UV detector.

L3 solution is the only **linearity** solution that has a Uracil peak present. This peak is used as a pump flow stability marker and has no effect on caffeine or **linearity** tests.