



## Which solution from the HSQ Kit do I use for injector carry over testing - FAQ

You should use the **L6 solution** which is the highest concentration of caffeine, 0.3490 mg/mL. This solution is the most likely to cause carry over.

Click [HERE](#) for HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024