



Which solution from the HSQ Kit do I use for injector carry over testing - FAQ

You should use the **L6 solution** which is the highest concentration of caffeine, 0.3490 mg/mL. This solution is the most likely to cause carry over.

Click [HERE](#) for HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-05-2024