
What is the acceptable wavelength band width for HPLC detectors when performing an HPLC Performance Qualification?

3nm is the normal setting for wavelength bandwidth for HPLC Detectors. 5nm may be considered too wide and 1nm is too narrow for PQ.